
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004686
11:04 08/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:29
PRESSURIZING 1:32
STERILIZING 2:34
136.5 C 221kPa 2:34
136.3 C 223kPa 3:04
136.1 C 222kPa 3:34
136.3 C 223kPa 4:04
136.3 C 222kPa 4:34
136.2 C 222kPa 5:04
136.4 C 223kPa 5:34
136.4 C 224kPa 6:04
136.4 C 224kPa 6:04
VENTING 6:04
AIR DRYING 6:50
CYCLE COMPLETE 13:22

Digital Signature #
B016B175D7DA52D5
