
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001894
7:51 26/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:36
STERILIZING 3:45
136.3 C 223kPa 3:45
136.3 C 224kPa 4:15
136.3 C 224kPa 4:45
136.2 C 222kPa 5:15
136.4 C 224kPa 5:45
136.4 C 225kPa 6:15
136.4 C 224kPa 6:45
136.0 C 220kPa 7:15
136.0 C 220kPa 7:15
VENTING 7:15
AIR DRYING 8:06
CYCLE COMPLETE 38:06

Digital Signature #
F0ED395FD00CA771
