
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000646
14:23 10/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:51
STERILIZING 4:05
136.0 C 228kPa 4:05
136.0 C 229kPa 4:35
136.3 C 231kPa 5:05
136.4 C 233kPa 5:35
136.4 C 233kPa 6:05
136.3 C 231kPa 6:35
136.3 C 233kPa 7:05
136.4 C 233kPa 7:35
136.4 C 233kPa 7:35
VENTING 7:35
AIR DRYING 8:29
CYCLE COMPLETE 23:29

Digital Signature #
EC8ECD2D2912C618
