
Midmark M11 - v1.0.5
Total Cycles: 3837

Name: RITTER2
Sterilizer ID: V2313228
08/07/2024 06:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.8 F 0.0
2:00 168.4 F 3.0
4:00 195.8 F 7.5
6:00 211.8 F 12.6
8:00 226.8 F 7.6
10:00 236.1 F 12.9
12:00 238.8 F 10.8
14:00 237.3 F 9.7
16:00 251.7 F 16.6
18:00 263.4 F 23.4
20:00 271.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.3 F 29.1
2:00 272.5 F 29.6
3:00 271.2 F 29.1
4:00 271.9 F 29.2

Min 271.1 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.8 F 0.0
5:00 221.9 F 0.0
10:00 193.7 F 0.0
15:00 188.8 F 0.0
20:00 184.7 F 0.0
25:00 181.7 F 0.0
30:00 180.9 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 20:03
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 00:59:33
CYCLE COMPLETE

08/07/2024 07:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____