
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002877
10:24 20/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:46
PRESSURIZING 1:49
STERILIZING 2:42
136.0 C 220kPa 2:42
136.2 C 224kPa 3:12
136.4 C 223kPa 3:42
136.4 C 224kPa 4:12
136.4 C 224kPa 4:42
136.4 C 222kPa 5:12
136.2 C 222kPa 5:42
136.3 C 224kPa 6:12
136.3 C 224kPa 6:12
VENTING 6:13
AIR DRYING 7:06
CYCLE COMPLETE 7:17

Digital Signature #
A772954E2E282779
