
Midmark M11 - v1.0.5
Total Cycles: 2921

Name: RITTER2
Sterilizer ID: V2313228
09/25/2023 02:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.6 F 0.0
2:00 167.3 F 3.0
4:00 192.4 F 7.3
6:00 210.2 F 11.8
8:00 221.6 F 6.5
10:00 234.6 F 11.7
12:00 235.2 F 9.3
14:00 239.2 F 10.0
16:00 247.8 F 14.3
18:00 259.3 F 20.6
20:00 269.4 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.2 F 29.3
2:00 273.0 F 29.6
3:00 271.6 F 29.6
4:00 271.7 F 29.3

Min 271.2 F 29.1
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 154.5 F 0.0
5:00 210.1 F 0.0
10:00 199.1 F 0.0
15:00 197.7 F 0.0
20:00 195.1 F 0.0
25:00 215.3 F 0.0
30:00 215.9 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 21:43
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:01:14
CYCLE COMPLETE

09/25/2023 03:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____