
Midmark M11 - v1.0.5
Total Cycles: 1996

Name: RITTER2
Sterilizer ID: V2313228
02/01/2023 04:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.7 F 0.0
2:00 177.3 F 4.0
4:00 190.2 F 8.7
6:00 213.3 F 8.6
8:00 223.9 F 9.2
10:00 233.7 F 9.3
12:00 242.5 F 13.0
14:00 242.8 F 12.2
16:00 256.1 F 19.2
18:00 267.8 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.3 F 29.5
2:00 271.6 F 29.4
3:00 272.2 F 29.4
4:00 271.4 F 29.2

Min 271.4 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.3 F 0.0
5:00 216.1 F 0.0
10:00 227.8 F 0.0
15:00 231.2 F 0.0
20:00 228.1 F 0.0
25:00 226.0 F 0.0
30:00 225.2 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 19:17
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:16
CYCLE COMPLETE

02/01/2023 05:01 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____