
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001313
16:26 20/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:47
STERILIZING 3:50
136.0 C 220kPa 3:50
136.2 C 223kPa 4:20
136.5 C 226kPa 4:50
136.2 C 224kPa 5:20
136.4 C 226kPa 5:50
136.4 C 225kPa 6:20
136.1 C 223kPa 6:50
136.1 C 223kPa 7:20
136.1 C 223kPa 7:20
VENTING 7:20
AIR DRYING 8:09
CYCLE COMPLETE 23:09

Digital Signature #
6875E16A9B333677
