
Midmark M11 - v1.0.0
Total Cycles: 3288

Name: RITTER 1
Sterilizer ID: V2178198
05/26/2023 08:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.4 F 0.0
2:00 79.8 F 0.0
4:00 87.5 F 0.0
6:00 116.4 F 0.0
8:00 164.4 F 0.0
10:00 186.7 F 1.5
12:00 203.2 F 5.1
14:00 216.7 F 8.3
16:00 230.4 F 11.7
18:00 228.7 F 6.0
20:00 236.2 F 10.6
22:00 238.0 F 7.9
24:00 245.1 F 13.1
26:00 242.8 F 11.5
28:00 255.2 F 17.8
30:00 266.0 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.7 F 28.8
2:00 271.7 F 28.0
3:00 272.2 F 28.7
4:00 272.4 F 28.5

Min 271.3 F 27.9
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.8 F 0.0
5:00 208.9 F 0.0
10:00 224.5 F 0.0
15:00 218.8 F 0.0
20:00 217.8 F 0.0
25:00 219.6 F 0.0
30:00 216.8 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:45
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:09:45
CYCLE COMPLETE

05/26/2023 09:54 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____