

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001063  
13:40 06/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:42  
PRESSURIZING 2:55  
STERILIZING 4:01  
136.0 C 220kPa 4:01  
136.4 C 224kPa 4:31  
136.0 C 222kPa 5:01  
136.4 C 225kPa 5:31  
136.2 C 222kPa 6:01  
136.4 C 225kPa 6:31  
136.2 C 222kPa 7:01  
136.2 C 223kPa 7:31  
136.2 C 223kPa 7:31  
VENTING 7:31  
AIR DRYING 8:20  
CYCLE COMPLETE 18:25  
-----

Digital Signature #  
8458BF0A2A59A539  
-----