

-----  
Midmark M11 - v1.0.5  
Total Cycles: 256

Name: RITTER2  
Sterilizer ID: V2313228  
07/28/2021 08:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 157.8 F 0.0  
2:00 161.1 F 2.3  
4:00 177.1 F 6.1  
6:00 200.4 F 10.6  
8:00 213.2 F 5.1  
10:00 226.7 F 9.6  
12:00 234.4 F 9.6  
14:00 241.3 F 12.5  
16:00 241.2 F 11.5  
18:00 254.4 F 18.1  
20:00 265.5 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.6 F 29.8  
2:00 271.7 F 29.1  
3:00 271.8 F 29.6  
4:00 272.4 F 29.7  
  
Min 271.2 F 28.9  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.8 F 0.0  
5:00 215.2 F 0.0  
10:00 207.6 F 0.0  
15:00 201.7 F 0.0  
20:00 192.0 F 0.0  
25:00 190.6 F 0.0  
30:00 171.1 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 21:48  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:00:22  
CYCLE COMPLETE

07/28/2021 09:58 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01