
Midmark M11 - v1.0.5
Total Cycles: 4228

Name: RITTER2
Sterilizer ID: V2313228
12/18/2024 04:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.3 F 0.0
2:00 169.5 F 2.8
4:00 186.3 F 7.0
6:00 207.2 F 12.2
8:00 225.4 F 6.9
10:00 238.1 F 12.1
12:00 236.1 F 9.6
14:00 237.6 F 9.2
16:00 248.2 F 14.8
18:00 260.0 F 21.4
20:00 270.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.6 F 29.4
2:00 271.5 F 29.1
3:00 272.3 F 29.4
4:00 271.5 F 28.8

Min 271.4 F 28.8
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.0 F 0.0
5:00 222.9 F 0.0
10:00 194.1 F 0.0
15:00 194.7 F 0.0
20:00 193.9 F 0.0
25:00 193.7 F 0.0
30:00 192.6 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 20:44
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 00:59:53
CYCLE COMPLETE

12/18/2024 05:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____