

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2226

Name: RITTER 1  
Sterilizer ID: V2178198  
08/16/2022 08:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.3 F 0.0  
2:00 150.7 F 0.9  
4:00 148.5 F 4.1  
6:00 173.2 F 8.2  
8:00 197.6 F 13.4  
10:00 215.2 F 5.8  
12:00 227.4 F 10.3  
14:00 230.9 F 7.8  
16:00 240.9 F 11.4  
18:00 237.1 F 9.0  
20:00 250.4 F 15.4  
22:00 261.5 F 21.7  
24:00 272.1 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.3  
1:00 272.0 F 28.7  
2:00 272.6 F 28.9  
3:00 271.7 F 28.3  
4:00 271.8 F 28.6

Min 271.3 F 28.1  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.3 F 0.0  
5:00 201.5 F 0.0  
10:00 182.5 F 0.0  
15:00 186.2 F 0.0  
20:00 189.8 F 0.0  
25:00 210.6 F 0.0  
30:00 203.0 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 24:27  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:02:39  
CYCLE COMPLETE

08/16/2022 09:30 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_