
Midmark M11 - v1.0.5
Total Cycles: 716

Name: RITTER2
Sterilizer ID: V2313228
01/20/2022 02:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.8 F 0.0
2:00 169.6 F 3.9
4:00 189.6 F 8.6
6:00 207.8 F 10.2
8:00 225.2 F 9.2
10:00 233.5 F 9.4
12:00 242.8 F 13.1
14:00 243.3 F 12.3
16:00 256.3 F 19.2
18:00 267.7 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.6 F 29.7
2:00 271.5 F 29.0
3:00 271.6 F 29.5
4:00 272.4 F 29.7

Min 271.0 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.6 F 0.0
5:00 208.8 F 0.0
10:00 215.6 F 0.0
15:00 201.6 F 0.0
20:00 199.4 F 0.0
25:00 197.0 F 0.0
30:00 204.2 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 19:22
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:19
CYCLE COMPLETE

01/20/2022 03:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____