
Midmark M11 - v1.0.0
Total Cycles: 2574

Name: RITTER 1
Sterilizer ID: V2178198
11/17/2022 10:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.8 F 0.0
2:00 193.6 F 3.1
4:00 208.1 F 7.7
6:00 223.8 F 12.9
8:00 226.9 F 7.2
10:00 239.6 F 12.4
12:00 234.1 F 9.0
14:00 237.9 F 9.9
16:00 245.6 F 13.4
18:00 257.8 F 20.0
20:00 268.7 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.0
1:00 272.2 F 29.5
2:00 272.7 F 29.4
3:00 271.8 F 28.8
4:00 272.6 F 29.8

Min 271.6 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.4 F 0.0
5:00 234.8 F 0.0
10:00 229.6 F 0.0
15:00 228.4 F 0.0
20:00 225.3 F 0.0
25:00 226.9 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 21:02
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:04
CYCLE COMPLETE

11/17/2022 10:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____