
Midmark M11 - v1.0.5
Total Cycles: 4453

Name: RITTER2
Sterilizer ID: V2313228
03/25/2025 12:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.8 F 0.0
2:00 163.6 F 2.5
4:00 177.3 F 6.4
6:00 197.1 F 10.7
8:00 223.7 F 5.8
10:00 229.7 F 9.4
12:00 236.6 F 10.4
14:00 240.5 F 11.9
16:00 239.7 F 10.7
18:00 252.0 F 16.9
20:00 263.0 F 23.2
22:00 272.6 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.8
1:00 272.0 F 29.9
2:00 272.5 F 30.2
3:00 271.3 F 30.1
4:00 271.8 F 29.9

Min 271.1 F 29.7
Max 272.7 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.2 F 0.0
5:00 234.1 F 0.0
10:00 219.3 F 0.0
15:00 216.8 F 0.0
20:00 212.8 F 0.0
25:00 211.4 F 0.0
30:00 207.8 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 23:04
STERILIZING: 4:00
VENTING: 3:57
DRYING: 30:00
TOTAL CYCLE: 01:02:21
CYCLE COMPLETE

03/25/2025 01:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____