
Midmark M11 - v1.0.0
Total Cycles: 2449

Name: RITTER 1
Sterilizer ID: V2178198
10/18/2022 11:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.6 F 0.0
2:00 174.6 F 3.7
4:00 185.3 F 8.5
6:00 203.4 F 9.4
8:00 223.9 F 8.3
10:00 237.8 F 13.0
12:00 238.0 F 11.0
14:00 237.2 F 9.5
16:00 251.5 F 16.5
18:00 263.7 F 23.7
20:00 271.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.6 F 29.5
2:00 271.6 F 28.7
3:00 272.6 F 29.6
4:00 272.2 F 29.1

Min 271.5 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 203.9 F 0.0
10:00 213.5 F 0.0
15:00 216.2 F 0.0
20:00 217.6 F 0.0
25:00 217.6 F 0.0
Dry Cycle Aborted

FILLING: 1:59
HEATING: 20:00
STERILIZING: 4:00
VENTING: 3:16
DRYING: 29:42
TOTAL CYCLE: 00:58:57

10/18/2022 12:47 PM
ACCEPTED: X
REJECTED:
Operator: U02