
Midmark M11 - v1.0.5
Total Cycles: 81

Name: RITTER2
Sterilizer ID: V2313228
06/09/2021 09:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.0 F 0.0
2:00 153.7 F 1.9
4:00 156.6 F 4.8
6:00 177.6 F 8.8
8:00 200.3 F 13.4
10:00 214.3 F 6.7
12:00 228.4 F 11.7
14:00 230.6 F 9.0
16:00 239.0 F 10.6
18:00 245.5 F 14.1
20:00 258.3 F 20.7
22:00 268.5 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.8
1:00 272.9 F 30.3
2:00 272.3 F 30.0
3:00 273.0 F 30.2
4:00 272.2 F 29.9

Min 272.1 F 29.7
Max 273.4 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.4 F 0.0
5:00 233.3 F 0.0
10:00 213.8 F 0.0
15:00 214.9 F 0.0
20:00 206.7 F 0.0
25:00 200.1 F 0.0
30:00 198.4 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 23:12
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:01:27
CYCLE COMPLETE

06/09/2021 10:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____