
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006781
9:01 22/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:23
PRESSURIZING 3:23
STERILIZING 4:24
136.4 C 221kPa 4:24
135.9 C 220kPa 4:54
136.4 C 222kPa 5:24
136.4 C 222kPa 5:54
136.0 C 219kPa 6:24
136.4 C 219kPa 6:54
136.4 C 221kPa 7:24
136.4 C 220kPa 7:54
136.4 C 220kPa 7:54
VENTING 7:55
AIR DRYING 8:45
CYCLE COMPLETE 38:46

Digital Signature #
7344EDCE43B74C67
