

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3388

Name: RITTER2  
Sterilizer ID: V2313228  
02/15/2024 02:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 89.3 F 0.0  
2:00 90.1 F 0.0  
4:00 99.2 F 0.6  
6:00 131.4 F 3.0  
8:00 151.6 F 6.3  
10:00 177.2 F 10.3  
12:00 199.4 F 6.4  
14:00 213.4 F 8.0  
16:00 232.9 F 12.5  
18:00 233.6 F 9.3  
20:00 236.8 F 9.9  
22:00 244.0 F 13.3  
24:00 255.6 F 19.4  
26:00 266.7 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.6  
1:00 271.1 F 29.6  
2:00 271.0 F 29.5  
3:00 273.3 F 30.5  
4:00 272.5 F 30.6

Min 270.9 F 29.5  
Max 273.6 F 30.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.7 F 0.0  
5:00 237.1 F 0.0  
10:00 228.0 F 0.0  
15:00 214.5 F 0.0  
20:00 204.4 F 0.0  
25:00 203.9 F 0.0  
30:00 195.0 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 27:30  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:05:43  
CYCLE COMPLETE

02/15/2024 03:39 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_