

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 006401  
7:40 04/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:57  
PRESSURIZING 2:58  
STERILIZING 3:59  
136.4 C 222kPa 3:59  
136.3 C 223kPa 4:29  
136.4 C 223kPa 4:59  
136.4 C 223kPa 5:29  
136.2 C 220kPa 5:59  
136.2 C 219kPa 6:29  
136.4 C 224kPa 6:59  
136.4 C 222kPa 7:29  
136.4 C 222kPa 7:29  
VENTING 7:30  
AIR DRYING 8:20  
CYCLE COMPLETE 10:15

-----  
Digital Signature #  
04FE7E1A8C848805  
-----