
Midmark M11 - v1.0.0
Total Cycles: 1373

Name: RITTER 1
Sterilizer ID: V2178198
11/16/2021 12:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.8 F 0.0
2:00 78.2 F 0.0
4:00 115.0 F 0.0
6:00 156.4 F 0.0
8:00 174.9 F 0.1
10:00 194.2 F 0.3
12:00 208.4 F 2.3
14:00 218.5 F 5.2
16:00 226.5 F 7.9
18:00 235.3 F 11.3
20:00 230.8 F 6.7
22:00 241.0 F 11.5
24:00 236.8 F 8.8
26:00 241.3 F 10.3
28:00 247.5 F 13.7
30:00 259.1 F 19.9
32:00 269.7 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.7 F 28.7
2:00 271.6 F 27.8
3:00 272.4 F 28.7
4:00 272.4 F 28.3

Min 271.4 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.9 F 0.0
5:00 231.0 F 0.0
10:00 223.4 F 0.0
15:00 220.1 F 0.0
20:00 218.0 F 0.0
25:00 216.0 F 0.0
30:00 216.6 F 0.0

Drying Complete

FILLING: 1:22
HEATING: 33:00
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:11:33
CYCLE COMPLETE

11/16/2021 02:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____