
Midmark M11 - v1.0.5
Total Cycles: 3675

Name: RITTER2
Sterilizer ID: V2313228
06/07/2024 08:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.3 F 0.0
2:00 69.6 F 0.1
4:00 75.5 F 0.7
6:00 107.6 F 2.7
8:00 142.7 F 5.6
10:00 168.9 F 8.7
12:00 188.6 F 12.8
14:00 207.0 F 5.5
16:00 216.8 F 9.3
18:00 231.0 F 13.5
20:00 234.9 F 9.5
22:00 238.1 F 10.2
24:00 244.0 F 12.8
26:00 255.3 F 18.8
28:00 265.9 F 25.2
30:00 271.8 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 272.4 F 29.6
2:00 272.9 F 30.0
3:00 272.0 F 29.8
4:00 272.0 F 29.4

Min 271.5 F 29.3
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.2 F 0.0
5:00 228.9 F 0.0
10:00 188.6 F 0.0
15:00 178.5 F 0.0
20:00 175.5 F 0.0
25:00 172.5 F 0.0
30:00 171.3 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:30
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:08:34
CYCLE COMPLETE

06/07/2024 09:30 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____