
Midmark M11 - v1.0.0
Total Cycles: 2275

Name: RITTER 1
Sterilizer ID: V2178198
08/31/2022 02:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.2 F 0.0
2:00 200.2 F 0.6
4:00 207.7 F 4.4
6:00 217.3 F 9.8
8:00 228.0 F 7.6
10:00 238.8 F 11.4
12:00 236.1 F 9.2
14:00 238.9 F 9.3
16:00 249.4 F 14.9
18:00 261.7 F 21.7
20:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.3 F 28.7
2:00 271.2 F 28.1
3:00 272.4 F 28.8
4:00 271.5 F 28.1

Min 271.2 F 28.0
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 231.9 F 0.0
10:00 237.9 F 0.0
15:00 233.1 F 0.0
20:00 228.6 F 0.0
25:00 229.2 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 20:25
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:58
CYCLE COMPLETE

08/31/2022 03:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____