
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000784
10:05 29/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:03
PRESSURIZING 2:08
STERILIZING 3:13
136.0 C 229kPa 3:13
136.4 C 233kPa 3:43
136.4 C 233kPa 4:13
136.2 C 231kPa 4:43
136.4 C 233kPa 5:13
136.3 C 232kPa 5:43
136.3 C 233kPa 6:13
136.1 C 233kPa 6:43
136.1 C 233kPa 6:43
VENTING 6:43
AIR DRYING 7:35
CYCLE COMPLETE 7:37

Digital Signature #
CE5E52214A222F9A
