
Midmark M11 - v1.0.0
Total Cycles: 3654

Name: RITTER 1
Sterilizer ID: V2178198
08/15/2023 11:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.8 F 0.0
2:00 189.6 F 3.7
4:00 206.8 F 9.2
6:00 228.2 F 7.4
8:00 233.7 F 10.2
10:00 234.0 F 7.9
12:00 244.6 F 13.4
14:00 243.3 F 12.0
16:00 255.8 F 18.5
18:00 266.7 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 271.8 F 28.4
2:00 272.4 F 28.8
3:00 272.5 F 29.1
4:00 271.6 F 28.6

Min 271.4 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.8 F 0.0
5:00 199.5 F 0.0
10:00 209.0 F 0.0
15:00 209.6 F 0.0
20:00 205.9 F 0.0
25:00 204.3 F 0.0
30:00 202.3 F 0.0

Drying Complete

FILLING: 2:06
HEATING: 19:28
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 00:58:57
CYCLE COMPLETE

08/15/2023 12:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____