
Midmark M11 - v1.0.0
Total Cycles: 2942

Name: RITTER 1
Sterilizer ID: V2178198
02/23/2023 05:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.6 F 0.0
2:00 184.1 F 4.7
4:00 200.7 F 9.6
6:00 218.7 F 6.3
8:00 229.2 F 10.0
10:00 232.8 F 8.0
12:00 244.6 F 13.4
14:00 244.1 F 12.4
16:00 257.4 F 19.4
18:00 269.1 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.4 F 29.0
2:00 271.4 F 28.4
3:00 272.5 F 29.2
4:00 271.8 F 28.6

Min 271.4 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.6 F 0.0
5:00 219.4 F 0.0
10:00 224.9 F 0.0
15:00 223.8 F 0.0
20:00 226.9 F 0.0
25:00 230.4 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 1:55
HEATING: 19:02
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:14
CYCLE COMPLETE

02/23/2023 06:14 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____