
Midmark M11 - v1.0.0
Total Cycles: 2674

Name: RITTER 1
Sterilizer ID: V2178198
12/14/2022 11:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.7 F 0.0
2:00 193.8 F 3.6
4:00 205.7 F 8.0
6:00 216.6 F 12.3
8:00 228.6 F 8.2
10:00 240.9 F 12.2
12:00 240.9 F 11.4
14:00 239.4 F 10.0
16:00 253.3 F 17.2
18:00 265.7 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 29.0
2:00 271.6 F 28.3
3:00 272.4 F 29.2
4:00 272.2 F 28.7

Min 271.5 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.8 F 0.0
5:00 227.0 F 0.0
10:00 233.8 F 0.0
15:00 229.1 F 0.0
20:00 224.2 F 0.0
25:00 224.3 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 19:39
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:01
CYCLE COMPLETE

12/14/2022 12:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____