
Midmark M11 - v1.0.5
Total Cycles: 4268

Name: RITTER2
Sterilizer ID: V2313228
01/15/2025 05:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.7 F 0.0
2:00 109.7 F 0.0
4:00 128.7 F 0.1
6:00 158.3 F 0.5
8:00 185.9 F 2.5
10:00 207.2 F 6.4
12:00 221.1 F 10.8
14:00 228.3 F 6.6
16:00 238.6 F 11.6
18:00 236.2 F 9.2
20:00 239.6 F 10.0
22:00 247.6 F 14.2
24:00 259.3 F 20.7
26:00 270.0 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.2 F 29.0
2:00 272.3 F 29.5
3:00 271.6 F 28.6
4:00 272.4 F 29.2

Min 271.3 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.1 F 0.0
5:00 195.2 F 0.0
10:00 191.6 F 0.0
15:00 192.2 F 0.0
20:00 191.1 F 0.0
25:00 189.9 F 0.0
30:00 188.5 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 26:53
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:05:20
CYCLE COMPLETE

01/15/2025 06:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____