

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000354  
15:26 19/08/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:54  
PRESSURIZING 3:00  
STERILIZING 3:57  
136.0 C 220kPa 3:57  
136.2 C 224kPa 4:27  
136.1 C 221kPa 4:57  
136.2 C 222kPa 5:27  
136.3 C 221kPa 5:57  
136.4 C 224kPa 6:27  
136.2 C 223kPa 6:57  
136.4 C 225kPa 7:27  
136.4 C 225kPa 7:27  
VENTING 7:27  
AIR DRYING 8:19  
CYCLE COMPLETE 8:46

-----  
Digital Signature #  
9855EB5763516C6C  
-----