
Midmark M11 - v1.0.5
Total Cycles: 3600

Name: RITTER2
Sterilizer ID: V2313228
05/08/2024 05:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 102.0 F 0.0
2:00 103.0 F 0.2
4:00 111.1 F 1.9
6:00 137.2 F 4.6
8:00 159.3 F 7.8
10:00 179.4 F 11.4
12:00 213.4 F 4.1
14:00 221.8 F 8.5
16:00 233.0 F 12.7
18:00 231.6 F 9.2
20:00 238.8 F 10.8
22:00 243.4 F 12.8
24:00 254.8 F 18.6
26:00 265.1 F 24.8
28:00 270.9 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 30.0
1:00 272.2 F 30.2
2:00 272.6 F 29.9
3:00 271.7 F 30.2
4:00 272.4 F 30.3

Min 271.4 F 29.9
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.5 F 0.0
5:00 235.0 F 0.0
10:00 177.3 F 0.0
15:00 174.3 F 0.0
20:00 173.7 F 0.0
25:00 173.6 F 0.0
30:00 174.4 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 28:42
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:07:18
CYCLE COMPLETE

05/08/2024 06:27 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____