
Midmark M11 - v1.0.5
Total Cycles: 3930

Name: RITTER2
Sterilizer ID: V2313228
09/11/2024 01:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.6 F 0.0
2:00 170.6 F 2.7
4:00 175.9 F 6.6
6:00 194.6 F 11.2
8:00 217.3 F 5.3
10:00 228.1 F 10.2
12:00 232.9 F 8.4
14:00 241.7 F 12.9
16:00 241.1 F 11.7
18:00 253.4 F 17.9
20:00 264.4 F 24.4
22:00 271.0 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.2
1:00 272.3 F 29.8
2:00 271.4 F 29.1
3:00 271.5 F 29.7
4:00 272.7 F 30.2

Min 270.9 F 29.0
Max 273.3 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 214.1 F 0.0
10:00 216.0 F 0.0
15:00 197.9 F 0.0
20:00 194.6 F 0.0
25:00 191.4 F 0.0
30:00 196.2 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 22:01
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:00:58
CYCLE COMPLETE

09/11/2024 02:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____