

-----  
Surrey, BC  
Canada  
-----

STATIM 5000 S501R711  
SN 511018F00076  
Unit # : 000  
WATER QUALITY  
2.3uS / 1.4ppm  
CYCLE NUMBER 000017  
15:44 13/01/2020  
HEAVY DUTY UNWRAPPED  
134 C for 6 min.  
CYCLE START 0:00  
CONDITIONING 1:57  
PRESSURIZING 5:02  
STERILIZING 6:39  
136.4 C 226kPa 6:39  
136.5 C 227kPa 7:10  
136.3 C 225kPa 7:40  
136.4 C 227kPa 8:10  
136.5 C 228kPa 8:40  
136.2 C 226kPa 9:10  
136.4 C 227kPa 9:40  
136.4 C 226kPa 10:10  
136.5 C 225kPa 10:40  
136.4 C 225kPa 11:10  
136.5 C 228kPa 11:40  
136.5 C 227kPa 12:10  
136.3 C 226kPa 12:40  
VENTING 12:40  
AIR DRYING 13:55  
CYCLE COMPLETE 15:56  
-----

Digital Signature #  
2DB55B05D7299724  
-----