
Midmark M11 - v1.0.5
Total Cycles: 1648

Name: RITTER2
Sterilizer ID: V2313228
10/24/2022 04:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.4 F 0.0
2:00 193.9 F 3.9
4:00 203.4 F 9.4
6:00 218.6 F 6.7
8:00 232.8 F 10.9
10:00 234.0 F 8.6
12:00 239.7 F 10.2
14:00 248.2 F 14.7
16:00 261.2 F 22.1
18:00 272.8 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.6 F 29.4
2:00 272.1 F 29.4
3:00 272.4 F 29.2
4:00 272.0 F 29.3

Min 271.6 F 28.8
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.5 F 0.0
5:00 213.8 F 0.0
10:00 232.6 F 0.0
15:00 231.5 F 0.0
20:00 223.1 F 0.0
25:00 218.9 F 0.0
30:00 216.3 F 0.0
Drying Complete

FILLING: 2:41
HEATING: 18:21
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:58:04
CYCLE COMPLETE

10/24/2022 05:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____