
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002514
8:37 19/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:10
STERILIZING 4:15
136.0 C 220kPa 4:15
135.9 C 220kPa 4:45
136.3 C 223kPa 5:15
136.3 C 222kPa 5:45
136.2 C 223kPa 6:15
136.4 C 224kPa 6:45
136.1 C 220kPa 7:15
136.4 C 225kPa 7:45
136.4 C 225kPa 7:45
VENTING 7:45
AIR DRYING 8:37
CYCLE COMPLETE 18:08

Digital Signature #
79C4322B6FA9D9F7
