

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2019

Name: RITTER2  
Sterilizer ID: V2313228  
02/07/2023 01:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 141.3 F 0.0  
2:00 142.4 F 0.0  
4:00 174.6 F 0.9  
6:00 200.6 F 4.8  
8:00 216.8 F 9.5  
10:00 226.3 F 8.0  
12:00 233.4 F 9.8  
14:00 236.1 F 9.0  
16:00 243.2 F 11.9  
18:00 245.9 F 13.5  
20:00 259.0 F 20.7  
22:00 270.7 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.4 F 29.3  
2:00 271.8 F 29.3  
3:00 272.2 F 29.2  
4:00 271.7 F 29.2

Min 271.5 F 28.7  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.4 F 0.0  
5:00 214.7 F 0.0  
10:00 241.4 F 0.0  
15:00 228.8 F 0.0  
20:00 225.9 F 0.0  
25:00 223.9 F 0.0  
30:00 222.9 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 22:45  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:00:46  
CYCLE COMPLETE

02/07/2023 02:13 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_