
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000380
15:08 21/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 2:03
STERILIZING 3:02
136.0 C 220kPa 3:02
136.2 C 231kPa 3:32
136.4 C 233kPa 4:02
136.3 C 233kPa 4:32
136.4 C 233kPa 5:02
136.3 C 232kPa 5:32
136.3 C 233kPa 6:02
136.4 C 233kPa 6:32
136.4 C 233kPa 6:32
VENTING 6:32
AIR DRYING 7:23
CYCLE COMPLETE 11:41

Digital Signature #
B3A20F939B0EA1D5
