

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4420

Name: RITTER2  
Sterilizer ID: V2313228  
03/11/2025 12:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 109.1 F 0.0  
2:00 109.4 F 0.8  
4:00 117.6 F 2.6  
6:00 141.7 F 5.8  
8:00 176.2 F 9.6  
10:00 198.2 F 7.6  
12:00 209.7 F 6.7  
14:00 223.7 F 10.7  
16:00 229.3 F 7.7  
18:00 238.3 F 12.0  
20:00 237.5 F 10.0  
22:00 249.8 F 16.0  
24:00 260.4 F 22.2  
26:00 270.6 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 30.2  
1:00 272.1 F 30.5  
2:00 272.1 F 30.2  
3:00 272.0 F 30.5  
4:00 272.1 F 30.6

Min 271.8 F 30.1  
Max 272.5 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.4 F 0.0  
5:00 206.7 F 0.0  
10:00 211.6 F 0.0  
15:00 197.5 F 0.0  
20:00 181.4 F 0.0  
25:00 180.1 F 0.0  
30:00 179.9 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 27:28  
STERILIZING: 4:00  
VENTING: 3:46  
DRYING: 30:00  
TOTAL CYCLE: 01:06:10  
CYCLE COMPLETE

03/11/2025 01:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_