
Midmark M11 - v1.0.5
Total Cycles: 1183

Name: RITTER2
Sterilizer ID: V2313228
06/09/2022 12:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.8 F 0.0
2:00 156.2 F 2.2
4:00 168.4 F 6.1
6:00 191.1 F 11.6
8:00 225.6 F 6.5
10:00 235.1 F 11.9
12:00 235.3 F 9.0
14:00 239.2 F 10.1
16:00 247.2 F 14.2
18:00 259.5 F 21.0
20:00 270.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.5
1:00 271.9 F 29.5
2:00 271.9 F 29.4
3:00 272.2 F 29.4
4:00 272.7 F 29.6

Min 271.6 F 29.3
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.4 F 0.0
5:00 238.4 F 0.0
10:00 218.5 F 0.0
15:00 236.7 F 0.0
20:00 225.9 F 0.0
25:00 227.6 F 0.0
30:00 222.4 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 21:29
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:01
CYCLE COMPLETE

06/09/2022 01:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____