
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001431
12:44 14/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:11
STERILIZING 4:13
136.0 C 221kPa 4:13
136.0 C 222kPa 4:43
136.3 C 225kPa 5:13
136.2 C 223kPa 5:43
136.3 C 224kPa 6:13
136.3 C 223kPa 6:43
136.3 C 223kPa 7:13
136.3 C 225kPa 7:43
136.3 C 225kPa 7:43
VENTING 7:44
AIR DRYING 8:36
CYCLE COMPLETE 9:15

Digital Signature #
605C94D3B109AB4F
