
Midmark M11 - v1.0.5
Total Cycles: 913

Name: RITTER2
Sterilizer ID: V2313228
03/28/2022 09:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.6 F 0.0
2:00 192.4 F 1.8
4:00 210.0 F 7.7
6:00 226.2 F 12.1
8:00 236.4 F 10.7
10:00 235.6 F 9.0
12:00 238.0 F 9.5
14:00 250.5 F 15.8
16:00 263.1 F 23.2
18:00 271.2 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.7 F 29.5
2:00 271.2 F 28.8
3:00 272.5 F 29.7
4:00 272.1 F 29.3

Min 271.1 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.3 F 0.0
5:00 231.2 F 0.0
10:00 217.0 F 0.0
15:00 227.5 F 0.0
20:00 222.9 F 0.0
25:00 232.0 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 18:04
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:56:56
CYCLE COMPLETE

03/28/2022 10:50 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____