
Midmark M11 - v1.0.0
Total Cycles: 4248

Name: RITTER 1
Sterilizer ID: V2178198
01/30/2024 01:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.4 F 0.0
2:00 112.5 F 0.9
4:00 116.5 F 2.8
6:00 140.6 F 5.5
8:00 165.9 F 8.7
10:00 186.0 F 12.5
12:00 214.0 F 4.7
14:00 216.2 F 8.6
16:00 229.2 F 12.9
18:00 234.0 F 8.8
20:00 242.8 F 11.9
22:00 242.9 F 11.7
24:00 254.1 F 17.4
26:00 263.8 F 23.4
28:00 271.3 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 271.6 F 29.0
2:00 271.7 F 29.1
3:00 271.6 F 29.0
4:00 271.6 F 28.9

Min 271.5 F 28.8
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.8 F 0.0
5:00 239.7 F 0.0
10:00 221.5 F 0.0
15:00 221.4 F 0.0
20:00 219.0 F 0.0
25:00 222.5 F 0.0
30:00 230.9 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 28:45
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:07:27
CYCLE COMPLETE

01/30/2024 02:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____