

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3256

Name: RITTER 1  
Sterilizer ID: V2178198  
05/18/2023 11:46 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.6 F 0.0  
2:00 144.9 F 0.0  
4:00 165.9 F 2.1  
6:00 197.2 F 6.4  
8:00 215.1 F 11.6  
10:00 223.6 F 6.1  
12:00 233.7 F 11.2  
14:00 233.6 F 8.1  
16:00 242.4 F 11.0  
18:00 246.5 F 13.2  
20:00 259.2 F 20.0  
22:00 270.4 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.6  
1:00 273.0 F 29.3  
2:00 272.3 F 28.7  
3:00 272.1 F 28.9  
4:00 272.9 F 29.3

Min 271.7 F 28.5  
Max 273.1 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.9 F 0.0  
5:00 200.9 F 0.0  
10:00 233.8 F 0.0  
15:00 230.4 F 0.0  
20:00 235.2 F 0.0  
25:00 228.4 F 0.0  
30:00 224.1 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 23:41  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:01:48  
CYCLE COMPLETE

05/18/2023 12:47 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_