
Midmark M11 - v1.0.5
Total Cycles: 926

Name: RITTER2
Sterilizer ID: V2313228
03/30/2022 08:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.3 F 0.0
2:00 120.6 F 0.0
4:00 132.4 F 0.0
6:00 177.4 F 0.4
8:00 202.5 F 2.4
10:00 216.3 F 6.6
12:00 227.9 F 11.1
14:00 230.1 F 7.4
16:00 241.8 F 13.2
18:00 241.8 F 11.8
20:00 241.0 F 11.0
22:00 254.4 F 17.8
24:00 266.0 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.5 F 29.2
2:00 271.5 F 28.7
3:00 272.7 F 29.3
4:00 271.5 F 28.5

Min 271.4 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.4 F 0.0
5:00 226.8 F 0.0
10:00 221.7 F 0.0
15:00 219.3 F 0.0
20:00 224.2 F 0.0
25:00 225.2 F 0.0
30:00 230.2 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 25:40
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:03:38
CYCLE COMPLETE

03/30/2022 09:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____