

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002392  
9:28 17/12/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:28  
PRESSURIZING 2:32  
STERILIZING 3:40  
136.0 C 220kPa 3:40  
136.4 C 223kPa 4:10  
136.3 C 224kPa 4:40  
136.0 C 222kPa 5:10  
136.2 C 223kPa 5:40  
136.2 C 222kPa 6:10  
136.3 C 224kPa 6:40  
136.0 C 219kPa 7:10  
136.0 C 219kPa 7:10  
VENTING 7:10  
AIR DRYING 8:05  
CYCLE COMPLETE 38:05  
-----

Digital Signature #  
A1707B8212362D3D  
-----