
Midmark M11 - v1.0.0
Total Cycles: 2058

Name: RITTER 1
Sterilizer ID: V2178198
06/27/2022 08:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.3 F 0.0
2:00 198.9 F 2.3
4:00 197.3 F 6.5
6:00 220.4 F 11.1
8:00 223.2 F 5.4
10:00 233.3 F 10.3
12:00 236.6 F 7.9
14:00 242.8 F 12.5
16:00 240.3 F 10.6
18:00 252.8 F 16.8
20:00 263.7 F 23.2
22:00 272.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 29.3
2:00 272.2 F 28.7
3:00 271.4 F 28.4
4:00 272.4 F 29.3

Min 271.4 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 227.7 F 0.0
10:00 228.8 F 0.0
15:00 228.9 F 0.0
20:00 231.4 F 0.0
25:00 229.4 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 22:16
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:39
CYCLE COMPLETE

06/27/2022 09:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____