
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001075
15:05 12/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:35
PRESSURIZING 4:36
STERILIZING 5:45
136.0 C 220kPa 5:45
136.3 C 225kPa 6:15
136.3 C 222kPa 6:45
136.4 C 225kPa 7:15
136.2 C 224kPa 7:45
136.4 C 224kPa 8:15
136.4 C 224kPa 8:45
136.4 C 226kPa 9:15
136.4 C 226kPa 9:15
VENTING 9:15
AIR DRYING 10:12
CYCLE COMPLETE 13:23

Digital Signature #
47C0D6B7CE9AAB0D
