
Midmark M11 - v1.0.5
Total Cycles: 4238

Name: RITTER2
Sterilizer ID: V2313228
01/02/2025 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.0 F 0.0
2:00 134.6 F 0.1
4:00 142.2 F 1.2
6:00 153.9 F 4.4
8:00 176.5 F 8.2
10:00 197.3 F 12.3
12:00 214.3 F 6.5
14:00 227.3 F 11.0
16:00 229.6 F 7.2
18:00 241.0 F 13.0
20:00 240.7 F 11.5
22:00 252.6 F 17.5
24:00 263.5 F 24.0
26:00 272.6 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.9
1:00 272.0 F 29.9
2:00 272.0 F 29.8
3:00 272.2 F 29.9
4:00 272.5 F 30.1

Min 271.8 F 29.8
Max 272.8 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 233.1 F 0.0
10:00 229.2 F 0.0
15:00 229.1 F 0.0
20:00 227.5 F 0.0
25:00 231.9 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 26:55
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:05:34
CYCLE COMPLETE

01/02/2025 01:30 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____