
Midmark M11 - v1.0.5
Total Cycles: 2617

Name: RITTER2
Sterilizer ID: V2313228
07/06/2023 11:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.6 F 0.0
2:00 185.6 F 2.6
4:00 204.2 F 6.6
6:00 223.7 F 11.8
8:00 226.6 F 7.2
10:00 234.7 F 12.6
12:00 237.1 F 10.0
14:00 235.9 F 8.7
16:00 248.6 F 15.0
18:00 260.3 F 21.8
20:00 271.2 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.6 F 30.0
2:00 272.3 F 29.4
3:00 271.3 F 29.0
4:00 271.8 F 29.8

Min 271.2 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.5 F 0.0
5:00 215.3 F 0.0
10:00 236.8 F 0.0
15:00 222.1 F 0.0
20:00 215.7 F 0.0
25:00 216.3 F 0.0
30:00 214.8 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 20:45
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:16
CYCLE COMPLETE

07/06/2023 12:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____