
Midmark M11 - v1.0.5
Total Cycles: 2925

Name: RITTER2
Sterilizer ID: V2313228
09/25/2023 07:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.9 F 0.0
2:00 177.8 F 4.0
4:00 204.0 F 9.8
6:00 219.3 F 5.8
8:00 232.2 F 11.6
10:00 235.0 F 9.8
12:00 236.1 F 8.7
14:00 250.9 F 16.2
16:00 263.7 F 23.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.1 F 28.9
2:00 272.4 F 29.5
3:00 271.5 F 28.6
4:00 272.2 F 29.3

Min 271.2 F 28.6
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 228.5 F 0.0
10:00 226.8 F 0.0
15:00 227.8 F 0.0
20:00 232.0 F 0.0
25:00 229.4 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 2:56
HEATING: 17:56
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:07
CYCLE COMPLETE

09/25/2023 08:03 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____